

Test Certificate

OFI Technologie & Innovation GmbH
 Firmensitz: 1030 Wien, Franz-Grill-Straße 5, Objekt 213
 L 16.01 - 0 | office@ofi.at | www.ofi.at

with reference to OFI Test Report No. 1800937/4786/4A from 2018-06-15

Client: Semperit Technische Produkte GmbH
 Modecenterstraße 22
 1030 Vienna
 AUSTRIA

Order: Test of single use gloves
 for compliance with ÖNORM EN 455-1, -2, -3

Samples: Medical examination gloves Sempercure Nitrile Skin², NOF-035VB,
 Size XS-XL, nonsterile, nitrile, powder free

Receipt of Samples: 2018-05-03

Tests: Tests were carried out in the individual technical departments of OFI within the scope of competence of the authorised signatories according to OFI QM manual.

Test Results:


| Test | Test method | Date of testing | Result |
|---|--------------------------|----------------------------|-------------------|
| Examination gloves for single use - Testing for leak tightness, (Acceptance Quality Level: AQL 1.5) | ÖNORM EN 455-1 (2001) | 2018-06-12 - 2018-06-15 | Compliance |
| Examination gloves for single use - Determination of breaking force Determination of dimensions | ÖNORM EN 455-2 (2015) | 2018-05-04 - 2018-05-17 | Compliance |
| Examination gloves for single use - Determination of residual powder | ÖNORM EN 455-3 (2015) | 2018-05-04 - 2018-05-17 | Compliance |

The testing results relate only to the samples analysed and refer to the detailed information of OFI Test report No. 1800937/4786/4A.

Approved:

G. Eitenberger

DI Gabriele Eitenberger
 Director in charge,
 Pharma, Medical Devices and Hygiene



Date:

18.06.2018

This certificate comprises 1 page.

It may be made available to third parties, either free of charge or against payment, if the full wording and if OFI is expressly named as the author. All tests applied are subject to a quality assurance programme according to EN ISO/IEC 17025:2007.